

# Sound Level Calibrators

**94dB or combination 94dB/114dB, 1kHz Professional Calibrator for 0.5" or 1.0" Microphones**  
With durable die-cast, aluminum housing

## Features:

- Calibrates and verifies Sound Level Meter operation
- For use on Sound Level Meters with 0.5" or 1.0" microphones
- Total Harmonic Distortion (THD) <2% at 94dB
- Meets IEC standard 60942 1997-11
- Complete with two 9V batteries, carrying case, and screwdriver

## Model 407744

- Battery test position with battery status LED indicates if battery is good
- 1kHz sine wave at 94dB is generated to an accuracy of 5% (frequency) and ±0.8dB

## Model 407766

- LED indicator lights up when power is on
- Level position to select 94dB or 114dB
- 1kHz sine wave at 94dB/114dB is generated to an accuracy of 5% (frequency) and ±0.8dB
- 3 position switch (OFF/94dB/114dB)



Specifications	Range	Accuracy
Frequency	1000 Hz	±5%
Sound Pressure Level	94 dB (407744); 94dB/114dB (407766)	±0.5dB (94dB), ±1dB (114dB)
Distortion	<2% Total Harmonic Distortion (THD)	
Operating Temperature	32 to 122°F (0 to 50°C)	
Power Supply	Two Heavy duty, alkaline 9V Battery	
Power consumption	Approx. 10mA DC	
Dimensions	2.2" diameter x 5.6" long (50 x 127mm)	
Weight	0.75 lbs. (340g)	



**GLOBAL CONTROL C.A.**  
Av. Bolivar Norte C.C. Camoruco  
Piso 15 Ofic 15-8 Valencia  
Telf. 0241-8244157 /8212318  
email: [info@globalcontrol.com.ve](mailto:info@globalcontrol.com.ve)  
[www.globalcontrol.com.ve](http://www.globalcontrol.com.ve)

## Ordering Information:

- 407744 .....Sound Level Calibrator, 94dB  
407744-NIST .....Sound Level Calibrator, 94dB with NIST Cert.  
407766 .....Sound Level Calibrator, 94dB/114dB  
407766-NIST .....Sound Level Calibrator, 94dB/114dB with NIST Cert.

